



Certificate of Analysis

Sterile Pack Swab, Carton of 200, Catalog No. 220518

Physical Characteristics

Appearance: Clear; opalescent or light precipitate is acceptable
pH at 25°C: 7.00 ± 0.50
Fill: 10 mL ± 1.0 mL

Biological Performance

Lot has been tested with the following organisms:

Culture	ATCC™	Result
<i>Candida albicans</i>	10231	Viable
<i>Salmonella choleraesuis</i> subsp. <i>choleraesuis</i>	13311	Viable
<i>Staphylococcus aureus</i>	6538	Viable
<i>Escherichia coli</i>	25922	Viable

Sterilization

Irradiation exposure is certified to be no less than 25 kGy.

Results reported were obtained at time of release.

Microb. Quality Assurance

ATCC is a trademark of the American Type Culture Collection.
BD and BD logo are trademarks of Becton, Dickinson and Company. © 2002 BD.

Becton, Dickinson and Company
7 Loveton Circle
Sparks, MD 21152 USA

Number of lot will be printed in this space